



CSA INTERNATIONAL

Certificate of Compliance

Certificate: 1694197

Master Contract: 231092

Project: 1694197

Date Issued: 2005/11/10

Issued to: Comark Corp.
93 West Street
Medfield, MA 02052
USA

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Abdelkrim Kassou, Jr. Eng.

Authorized by: Helene Vaillancourt, Manager,
Certification Services

PRODUCTS

CLASS 8750 01 - MEDICAL ELECTRICAL EQUIPMENT

CLASS 8750 81 - MEDICAL ELECTRICAL EQUIPMENT - Certified to US Standards

Medical computer, ITX Medical Chassis, model 06-10175-001, rated 12Vdc, 6A, powered by an external power supply, rated: Input: 100-250 Vac, 47-63 Hz, 3A, Output: 12Vdc, 9A.

- Type of protection against electric shock: Class I
- Degree of protection against electric shock: Not Classified
- Degree of protection against ingress of water: IPX0
- Degree of safety of application in the presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide: Equipment not suitable for use in the presence of a flammable anesthetic mixture with air or

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1694197

Master Contract: 231092

Project: 1694197

Date Issued: 2005/11/10

-
- with oxygen or nitrous oxide.
• Mode of operation:

Continuous

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 601.1-M90 -Medical Electrical Equipment Part I: General Requirements for Safety

CAN/CSA-C22.2 No. 601.1S1-94 -Supplement No. 1-94 to CAN/CSA-C22.2 No. 601.1-M90--Medical Electrical Equipment--Part 1: General Requirements for Safety

CAN/CSA-C22.2 No. 601.1-98 -Medical Electrical Equipment Part I: General Requirements for Safety Amendment 2

UL Standard No 60601-1 -Medical Electrical Equipment Part I: General Requirements for Safety

Subject to the following conditions:

- (1) The equipment has not been investigated for the protection against hazards of explosions in medically used rooms.
- (2) Units provided with other than CSA Certified power supply cord sets are certified as components only.
- (3) Evaluated to IEC/CSA 601 1 Amendment 2 excluding requirements for Electromagnetic compatibility (Clause 36), Biocompatibility (Clause 48) and Programmable Electronic Systems (IEC 60601 1 4 referenced in sub clause 52.1).