

Certificate of Conformity

issued by Curtis-Straus LLC, a wholly owned subsidiary of BV CPS

Company: Foresight Imaging
Product Tested: TIMS DICOM System
Testing Date: September 17 to October 2, 2012
Report No.: EM2062-1

A sample of the product as configured in the accompanying test report has been found to comply with the following standards:

- EN 60601-1-2:2007 Electromagnetic emissions and immunity requirements for medical electrical equipment – Group 1 Equipment, Class A for non-life supporting equipment
- IEC 60601-1-2 Ed. 2.1 Electromagnetic emissions and immunity requirements for medical electrical equipment – Group 1 Equipment, Class A for non-life supporting equipment

Attested to by the hands and seals:

For Curtis-Straus LLC

For the Manufacturer or Importer

Disha Vachhani 11/7/12
Test Engineer Date

[Signature] 11/7/12
Date

[Signature] 11/7/12
EMC Manager Date